Membrane Solutions
Tel:+1-866-528-4572 **Fax:+** 1-732-412-4040
E-mail :info@membrane-solutions.com
www.membrane-solutions.com

Certificate of Analysis

Hydrophobic PTFE Membrane for Venting,Pore Size:0.1µm, Width:270mm, Support Layer:/	**Part Number: FPB010N**
	Lot Number: 321059082

ISO 9001 Quality Standard
The product was manufactured under Membrane Solutions Quality Management System and is approved by an accredited registering body to the ISO 9001 Quality Systems Standard.

Membrane Filter Lot Characterization			
Filter Media	PTFE	**Wettability**	Hydrophobic
Pore Size(µ m)	0.10	**Thickness(µm)**	60-75
Bubble Point (Mpa@60%IPA)	0.308-0.323	**Air Permeability(L/ min/cm²@7KPa)**	0.25-0.31

We hereby certify that the above stated filters are in compliance with all applicable instructions. This certificate shall not be copied or reproduced without the written approval of Membrane Solutions corp.

Tested By:*Tommy* 5-8-2024

R & D Engineer